

SAFECHEM Europe GmbH
Georg-Glock-Str. 3
40474 Düsseldorf

Proof of Conformity

In-vitro-cytotoxicity according to DIN ISO 10993-5, 2009

Test Unit	Prüfbereich ZS Fraunhofer-Institut für Grenzflächen- und Bioverfahrenstechnik (IGB) Nobelstrasse 12 70569 Stuttgart
Company	SAFECHEM Europe GmbH
Offer Number	11 14 001
Test method	<i>In-vitro</i> -cytotoxicity according to DIN ISO 10993-5, 2009
Period of testing	16.12. 2013 – 14.02.2014
Test item	metal plates of titanium alloy, Grade 5Ti-6Al-4V, cleaned with the solvent DOWCLENETM 1601
Internal code	ZS / 1 / 11 14 001 / 1
Extraction	conform with DIN ISO 10993-12 72h ± 2h at 37°C ± 1°C and 5% CO ₂

Result of extract testing

The extract of the metal plates of titanium alloy, Grade 5 Ti-6Al-4V, cleaned with the solvent DOWCLENETM 1601 was tested non cytotoxic.

18.02.2014 Dipl.-Biol (f.o.) J. Thiele
Date, Title and Signature of Testing Leader